

Quality Certificate
Technical Data for Validation of Free
Ammonia Chemkey® Reagents

Quality Certificate

Technical Data for Free Ammonia Chemkeys:

Part Number	9429500
Sensitivity	0.874 Abs / mg/L
Ordinate Intersect	0.004 mg/L
Standard Deviation	< 0.04 mg/L at 0.30 mg/L
Accuracy	< ± 8% at 0.30 mg/L
Correlation Coefficient (r ²)	> 0.990

The detection and the quantitation limits were determined in conformity with USEPA method 40 CFR Appendix B to Part 136, Revision 1.11.

Range (mg/L as NH ₃ -N)	0.05 – 0.50
Detection limit	< 0.05 mg/L

The series of the smallest and largest calibration standards exhibit normal distribution and are outlier- and trend-free.

The calibration gives a linear function.

Results are certified for nine months after manufacture date when stored at 25°C or below.

Result	Precision Confidence Interval (95%)
0.30 mg/L	± 0.03 mg/L
0.50 mg/L	± 0.03 mg/L



Bennett Vandiver
Sr. Manager, Strategic Quality
Hach



Darren MacFarland, Ph.D.
Director, Chemistry R&D
Hach



Be Right™

HACH
5600 Lindbergh Drive
Loveland, CO USA, 80538
800-227-4224
www.hach.com

HACH
Willstätterstr 11, 40549
Düsseldorf, Germany
(+49) (211) 52880
www.hach-lange.com